

	ACCEPTABILITY CRITERIA FOR BIOLOGICAL SAMPLES	REVISION 01.25 year 2025 Pag. 1 a 1
---	--	---

Next Lab Italy laboratories may not accept the biological sample received and therefore request a repetition of the sampling or a new sample collection, for example if one of the following cases occurs:

1. Insufficient sample (i.e. due to low cellularity);
2. Hemolyzed sample;
3. Coagulated sample;
4. Samples not correctly identified;
5. Mismatch between the personal data reported on the test tube and/or container and the personal data reported on the request;
6. Type of test tubes and/or containers not suitable for performing the requested test (i.e. NIPT samples performed in non-STRECK tubes);
7. Samples with transport/storage temperatures not suitable for the requested test;
8. Samples arriving at the laboratory after the stability deadlines.

In the event that the above-mentioned suitability requirements are not met, depending on the type of test, the laboratory will notify the referring specialist either by telephone or by issuing a non-compliance, which will require resolution in writing. Depending on the non-compliance detected, a new biological sample may be requested for which the patient will have to undergo a new biological sampling.